
Midmark M11 - v1.0.0
Total Cycles: 2169

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2022 12:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.3 F 0.0
2:00 161.1 F 3.0
4:00 180.2 F 7.3
6:00 201.4 F 12.1
8:00 220.7 F 6.2
10:00 233.5 F 11.3
12:00 231.6 F 7.7
14:00 245.7 F 13.2
16:00 243.2 F 12.0
18:00 255.7 F 18.4
20:00 266.7 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.4 F 29.1
2:00 272.8 F 29.1
3:00 271.7 F 28.7
4:00 272.7 F 29.2

Min 271.0 F 28.1
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 233.2 F 0.0
10:00 227.7 F 0.0
15:00 226.8 F 0.0
20:00 223.0 F 0.0
25:00 228.2 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 21:37
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

07/25/2022 01:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____