

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4798

Name: RITTER 1  
Sterilizer ID: V2178198  
08/08/2024 03:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.4 F 0.0  
2:00 186.9 F 4.2  
4:00 195.1 F 9.1  
6:00 224.6 F 8.8  
8:00 227.1 F 8.9  
10:00 235.7 F 10.0  
12:00 240.2 F 11.8  
14:00 239.2 F 10.2  
16:00 252.7 F 17.0  
18:00 264.2 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 271.8 F 28.7  
2:00 272.3 F 29.0  
3:00 272.4 F 29.3  
4:00 271.4 F 28.7  
  
Min 271.4 F 28.6  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 200.6 F 0.0  
10:00 210.9 F 0.0  
15:00 208.3 F 0.0  
20:00 201.7 F 0.0  
25:00 202.4 F 0.0  
30:00 204.2 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 19:54  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 00:58:53  
CYCLE COMPLETE

08/08/2024 03:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_