
Midmark M11 - v1.0.5
Total Cycles: 304

Name: RITTER2
Sterilizer ID: V2313228
08/24/2021 06:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.8 F 0.0
2:00 79.4 F 0.0
4:00 76.7 F 0.0
6:00 109.1 F 0.2
8:00 142.7 F 2.0
10:00 173.4 F 5.0
12:00 192.2 F 8.5
14:00 202.2 F 12.9
16:00 215.3 F 6.6
18:00 226.2 F 10.7
20:00 230.8 F 7.9
22:00 241.3 F 12.4
24:00 240.0 F 10.8
26:00 252.5 F 17.0
28:00 263.0 F 23.2
30:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.3 F 29.6
2:00 271.5 F 28.8
3:00 272.0 F 29.7
4:00 272.3 F 29.3

Min 271.1 F 28.7
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 233.4 F 0.0
10:00 209.7 F 0.0
15:00 216.5 F 0.0
20:00 212.2 F 0.0
25:00 215.1 F 0.0
30:00 207.6 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 30:21
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:08:37
CYCLE COMPLETE

08/24/2021 07:57 AM
ACCEPTED: X
REJECTED:
Operator: U02