
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 007097
10:04 25/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:27
STERILIZING 3:25
136.4 C 224kPa 3:25
136.3 C 224kPa 3:55
136.4 C 222kPa 4:25
136.4 C 223kPa 4:55
136.3 C 225kPa 5:25
136.4 C 223kPa 5:55
136.4 C 225kPa 6:25
136.1 C 218kPa 6:55
136.1 C 218kPa 6:55
VENTING 6:55
AIR DRYING 7:50
CYCLE COMPLETE 7:54

Digital Signature #
04FA13525A130683
