

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3622

Name: RITTER2  
Sterilizer ID: V2313228  
05/15/2024 02:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.6 F 0.0  
2:00 171.5 F 3.3  
4:00 191.9 F 8.0  
6:00 211.1 F 12.9  
8:00 220.8 F 7.9  
10:00 234.1 F 13.3  
12:00 237.4 F 11.0  
14:00 236.8 F 9.7  
16:00 250.7 F 16.3  
18:00 262.2 F 22.9  
20:00 272.5 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.6 F 30.0  
2:00 271.9 F 29.4  
3:00 271.3 F 29.4  
4:00 272.5 F 30.1  
  
Min 271.2 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.5 F 0.0  
5:00 215.9 F 0.0  
10:00 204.8 F 0.0  
15:00 201.2 F 0.0  
20:00 197.8 F 0.0  
25:00 195.6 F 0.0  
30:00 199.2 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 20:29  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:00:01  
CYCLE COMPLETE

05/15/2024 03:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_