
Midmark M11 - v1.0.0
Total Cycles: 1637

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2022 01:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.0 F 0.0
2:00 177.5 F 2.9
4:00 187.5 F 7.0
6:00 206.1 F 12.3
8:00 228.0 F 7.2
10:00 234.1 F 12.6
12:00 238.4 F 9.8
14:00 237.4 F 8.5
16:00 249.8 F 14.7
18:00 261.9 F 21.6
20:00 272.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.5 F 28.9
2:00 272.8 F 28.8
3:00 271.7 F 28.2
4:00 272.4 F 28.9

Min 271.4 F 28.1
Max 273.1 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 232.7 F 0.0
10:00 229.1 F 0.0
15:00 230.1 F 0.0
20:00 229.9 F 0.0
25:00 230.6 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 21:02
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

02/15/2022 03:05 PM
ACCEPTED: X
REJECTED:
Operator: U02