
Midmark M11 - v1.0.0
Total Cycles: 5323

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2025 11:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.5 F 0.0
2:00 167.1 F 2.2
4:00 174.8 F 5.7
6:00 189.2 F 9.3
8:00 206.9 F 13.1
10:00 217.2 F 4.9
12:00 227.6 F 8.9
14:00 238.4 F 13.2
16:00 238.2 F 9.8
18:00 239.2 F 9.7
20:00 246.6 F 13.4
22:00 257.4 F 19.2
24:00 267.2 F 25.2
26:00 271.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.3 F 29.2
2:00 272.6 F 29.2
3:00 272.5 F 28.9
4:00 271.6 F 28.6

Min 271.3 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.2 F 0.0
5:00 233.0 F 0.0
10:00 203.7 F 0.0
15:00 193.1 F 0.0
20:00 189.5 F 0.0
25:00 187.9 F 0.0
30:00 186.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:42
STERILIZING: 4:00
VENTING: 4:02
DRYING: 30:00
TOTAL CYCLE: 01:05:38
CYCLE COMPLETE

02/07/2025 12:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____