
Midmark M11 - v1.0.0
Total Cycles: 2373

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2022 04:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.7 F 0.0
2:00 196.5 F 4.7
4:00 202.6 F 9.7
6:00 221.0 F 6.5
8:00 233.9 F 10.7
10:00 232.1 F 7.8
12:00 240.7 F 11.0
14:00 245.8 F 13.4
16:00 258.6 F 20.3
18:00 270.0 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.3 F 29.3
2:00 271.6 F 28.6
3:00 271.8 F 29.1
4:00 272.2 F 29.2

Min 271.2 F 28.4
Max 272.3 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 228.6 F 0.0
10:00 234.4 F 0.0
15:00 236.8 F 0.0
20:00 229.3 F 0.0
25:00 229.0 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:06
CYCLE COMPLETE

09/26/2022 05:57 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____