

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4461

Name: RITTER 1  
Sterilizer ID: V2178198  
04/16/2024 08:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.3 F 0.0  
2:00 71.8 F 0.3  
4:00 83.3 F 1.2  
6:00 119.7 F 3.3  
8:00 144.7 F 6.1  
10:00 173.3 F 9.6  
12:00 192.2 F 13.3  
14:00 210.5 F 3.9  
16:00 222.4 F 7.1  
18:00 233.1 F 10.7  
20:00 232.5 F 7.7  
22:00 239.6 F 10.7  
24:00 234.4 F 7.6  
26:00 247.5 F 13.8  
28:00 258.4 F 19.8  
30:00 268.7 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.2 F 28.4  
2:00 272.5 F 28.8  
3:00 272.6 F 29.1  
4:00 271.5 F 28.6

Min 271.4 F 28.3  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.1 F 0.0  
5:00 223.0 F 0.0  
10:00 214.1 F 0.0  
15:00 216.9 F 0.0  
20:00 205.4 F 0.0  
25:00 210.4 F 0.0  
30:00 210.9 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:52  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:10:04  
CYCLE COMPLETE

04/16/2024 09:27 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_