

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006412  
10:02 05/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:55  
PRESSURIZING 2:56  
STERILIZING 3:59  
136.4 C 223kPa 3:59  
136.1 C 222kPa 4:29  
136.4 C 225kPa 4:59  
136.4 C 224kPa 5:29  
136.4 C 224kPa 5:59  
136.4 C 224kPa 6:29  
136.4 C 224kPa 6:59  
136.4 C 221kPa 7:29  
136.4 C 221kPa 7:29  
VENTING 7:29  
AIR DRYING 8:22  
CYCLE COMPLETE 8:27

-----  
Digital Signature #  
B5A254046958876F  
-----