
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001564
15:48 06/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:15
STERILIZING 4:26
136.0 C 222kPa 4:26
136.4 C 225kPa 4:56
136.3 C 224kPa 5:26
136.1 C 223kPa 5:56
136.2 C 225kPa 6:26
136.0 C 222kPa 6:56
136.4 C 225kPa 7:26
136.4 C 225kPa 7:56
136.4 C 225kPa 7:56
VENTING 7:56
AIR DRYING 8:55
CYCLE COMPLETE 9:09

Digital Signature #
386B1E1D274B98B0
