

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005657  
11:30 02/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:50  
PRESSURIZING 1:58  
STERILIZING 3:06  
136.4 C 220kPa 3:06  
136.2 C 223kPa 3:36  
136.2 C 222kPa 4:06  
136.4 C 224kPa 4:36  
136.1 C 222kPa 5:06  
136.4 C 225kPa 5:36  
136.4 C 225kPa 6:06  
136.4 C 224kPa 6:36  
136.4 C 224kPa 6:36  
VENTING 6:36  
AIR DRYING 7:28  
CYCLE COMPLETE 9:26

-----  
Digital Signature #  
E4323CF48D258D61  
-----