
Midmark M11 - v1.0.5
Total Cycles: 4275

Name: RITTER2
Sterilizer ID: V2313228
01/17/2025 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.4 F 0.0
2:00 162.3 F 0.8
4:00 182.9 F 4.6
6:00 201.3 F 8.6
8:00 217.6 F 13.2
10:00 225.4 F 8.1
12:00 237.0 F 13.2
14:00 239.0 F 11.1
16:00 237.9 F 9.9
18:00 251.1 F 16.3
20:00 262.1 F 22.7
22:00 272.3 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.7
2:00 272.0 F 29.0
3:00 271.9 F 29.4
4:00 272.6 F 29.5

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 191.5 F 0.0
10:00 210.1 F 0.0
15:00 206.5 F 0.0
20:00 204.2 F 0.0
25:00 203.6 F 0.0
30:00 203.1 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 22:30
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:01:42
CYCLE COMPLETE

01/17/2025 01:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____