
Midmark M11 - v1.0.0
Total Cycles: 4741

Name: RITTER 1
Sterilizer ID: V2178198
07/17/2024 11:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.0 F 0.0
2:00 179.7 F 3.8
4:00 187.5 F 8.1
6:00 206.3 F 13.1
8:00 218.8 F 7.6
10:00 231.6 F 12.9
12:00 235.5 F 10.0
14:00 235.1 F 8.4
16:00 248.4 F 14.9
18:00 260.6 F 21.6
20:00 271.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.0
2:00 272.5 F 29.5
3:00 271.5 F 28.7
4:00 272.7 F 29.4

Min 271.3 F 28.7
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 217.9 F 0.0
10:00 216.0 F 0.0
15:00 227.9 F 0.0
20:00 236.0 F 0.0
25:00 226.4 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 00:59:30
CYCLE COMPLETE

07/17/2024 12:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____