
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 001464
12:47 09/30/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:02
PRESSURIZING 5:39
STERILIZING 7:03
136.0 C 220kPa 7:03
136.1 C 220kPa 7:33
136.1 C 218kPa 8:03
136.3 C 222kPa 8:33
136.0 C 221kPa 9:03
136.4 C 220kPa 9:33
136.3 C 222kPa 10:03
136.5 C 223kPa 10:33
VENTING 10:33
AIR DRYING 11:37
CYCLE COMPLETE 13:57

Digital Signature #
4B5FCAC75DCF69C9
