
Midmark M11 - v1.0.5
Total Cycles: 4327

Name: RITTER2
Sterilizer ID: V2313228
02/03/2025 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.7 F 0.0
2:00 134.2 F 1.0
4:00 136.3 F 3.7
6:00 160.8 F 7.3
8:00 189.9 F 11.5
10:00 210.6 F 4.8
12:00 224.3 F 9.0
14:00 233.0 F 13.5
16:00 234.5 F 9.9
18:00 237.5 F 9.4
20:00 246.3 F 13.9
22:00 257.9 F 20.3
24:00 268.5 F 27.2
26:00 271.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.8
1:00 272.7 F 30.3
2:00 271.7 F 29.9
3:00 272.7 F 30.3
4:00 271.8 F 29.7

Min 271.6 F 29.7
Max 273.1 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.9 F 0.0
5:00 234.5 F 0.0
10:00 214.7 F 0.0
15:00 205.0 F 0.0
20:00 192.9 F 0.0
25:00 192.3 F 0.0
30:00 187.1 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:03
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:04:31
CYCLE COMPLETE

02/03/2025 12:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____