
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000327
11:31 11/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:10
STERILIZING 3:39
136.5 C 233kPa 3:39
136.4 C 233kPa 4:09
136.3 C 232kPa 4:39
136.4 C 231kPa 5:09
136.2 C 231kPa 5:39
136.4 C 232kPa 6:09
136.4 C 233kPa 6:39
136.4 C 233kPa 7:09
136.4 C 233kPa 7:09
VENTING 7:09
AIR DRYING 8:04
CYCLE COMPLETE 23:04

Digital Signature #
B03024F3150CE778
