
Midmark M11 - v1.0.0
Total Cycles: 4532

Name: RITTER 1
Sterilizer ID: V2178198
05/07/2024 02:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.7 F 0.0
2:00 95.8 F 0.1
4:00 108.5 F 1.7
6:00 137.9 F 4.3
8:00 166.7 F 7.5
10:00 191.2 F 10.8
12:00 215.5 F 3.7
14:00 217.5 F 7.5
16:00 230.7 F 11.5
18:00 228.2 F 7.4
20:00 242.0 F 12.9
22:00 239.8 F 10.8
24:00 252.5 F 16.9
26:00 263.5 F 23.3
28:00 272.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.7 F 29.3
2:00 271.4 F 28.6
3:00 272.6 F 29.4
4:00 272.0 F 28.8

Min 271.2 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 210.5 F 0.0
10:00 211.2 F 0.0
15:00 212.8 F 0.0
20:00 216.8 F 0.0
25:00 219.5 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:14
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:06:34
CYCLE COMPLETE

05/07/2024 03:57 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____