

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000911  
11:11 07/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:37  
STERILIZING 3:46  
136.0 C 220kPa 3:46  
136.1 C 228kPa 4:16  
135.9 C 229kPa 4:46  
136.0 C 229kPa 5:16  
136.2 C 231kPa 5:46  
135.9 C 230kPa 6:16  
136.4 C 233kPa 6:46  
136.4 C 232kPa 7:16  
136.4 C 232kPa 7:16  
VENTING 7:16  
AIR DRYING 8:09  
CYCLE COMPLETE 23:09  
-----

Digital Signature #  
B47AF270F6D19A59  
-----