

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1343

Name: RITTER 1  
Sterilizer ID: V2178198  
11/09/2021 07:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.2 F 0.0  
2:00 75.8 F 0.0  
4:00 79.4 F 0.0  
6:00 106.1 F 0.0  
8:00 137.5 F 0.5  
10:00 165.5 F 2.4  
12:00 184.9 F 4.8  
14:00 200.9 F 7.9  
16:00 213.9 F 11.1  
18:00 221.6 F 5.9  
20:00 228.4 F 8.5  
22:00 238.5 F 12.8  
24:00 237.0 F 9.6  
26:00 238.6 F 9.6  
28:00 246.6 F 13.4  
30:00 257.5 F 19.3  
32:00 267.5 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.5 F 28.9  
2:00 271.8 F 28.2  
3:00 271.9 F 28.6  
4:00 272.4 F 28.8  
  
Min 271.4 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.8 F 0.0  
5:00 199.8 F 0.0  
10:00 205.2 F 0.0  
15:00 212.2 F 0.0  
20:00 213.6 F 0.0  
25:00 209.0 F 0.0  
30:00 194.3 F 0.0  
  
Drying Complete

FILLING: 1:01  
HEATING: 33:32  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:11:52  
CYCLE COMPLETE

11/09/2021 08:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_