
Midmark M11 - v1.0.5
Total Cycles: 139

Name: RITTER2
Sterilizer ID: V2313228
06/24/2021 12:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.2 F 0.0
2:00 169.6 F 3.1
4:00 185.4 F 6.8
6:00 204.4 F 11.4
8:00 217.7 F 5.9
10:00 228.5 F 10.9
12:00 231.8 F 7.9
14:00 240.9 F 12.1
16:00 244.3 F 13.2
18:00 256.1 F 19.5
20:00 266.8 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.5
1:00 272.1 F 29.4
2:00 272.3 F 29.7
3:00 272.4 F 30.1
4:00 271.7 F 29.5

Min 271.7 F 29.4
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.8 F 0.0
5:00 222.3 F 0.0
10:00 202.3 F 0.0
15:00 201.2 F 0.0
20:00 200.3 F 0.0
25:00 198.9 F 0.0
30:00 199.6 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 21:30
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

06/24/2021 01:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____