
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001127
14:37 20/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:16
STERILIZING 4:21
136.0 C 221kPa 4:21
136.2 C 223kPa 4:51
136.3 C 225kPa 5:21
136.4 C 225kPa 5:51
136.2 C 224kPa 6:21
136.0 C 222kPa 6:51
136.4 C 226kPa 7:21
136.2 C 223kPa 7:51
136.2 C 223kPa 7:51
VENTING 7:51
AIR DRYING 8:43
CYCLE COMPLETE 19:19

Digital Signature #
4C1D4040F7331826
