

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1586

Name: RITTER2  
Sterilizer ID: V2313228  
10/12/2022 01:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.2 F 0.0  
2:00 193.3 F 3.7  
4:00 200.8 F 8.6  
6:00 221.1 F 10.2  
8:00 230.5 F 10.1  
10:00 233.5 F 8.3  
12:00 242.0 F 12.0  
14:00 245.2 F 13.3  
16:00 257.7 F 20.1  
18:00 269.0 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.5 F 29.7  
2:00 271.4 F 28.9  
3:00 272.4 F 29.9  
4:00 271.9 F 29.3  
  
Min 271.3 F 28.9  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 229.7 F 0.0  
10:00 251.5 F 0.0  
15:00 239.2 F 0.0  
20:00 226.9 F 0.0  
25:00 222.3 F 0.0  
30:00 218.8 F 0.0  
Drying Complete

FILLING: 1:60  
HEATING: 19:07  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:15  
CYCLE COMPLETE

10/12/2022 02:42 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_