

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000656  
11:32 13/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:40  
PRESSURIZING 2:49  
STERILIZING 3:50  
136.0 C 227kPa 3:50  
136.1 C 230kPa 4:20  
136.3 C 231kPa 4:50  
136.2 C 231kPa 5:20  
136.3 C 231kPa 5:50  
136.4 C 233kPa 6:20  
136.4 C 232kPa 6:50  
136.4 C 231kPa 7:20  
136.4 C 231kPa 7:20  
VENTING 7:20  
AIR DRYING 8:15  
CYCLE COMPLETE 8:19  
-----

Digital Signature #  
7FDC14CC1D04794C  
-----