
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000346
7:18 20/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:21
CASSETTE REMOVED
NOT STERILE
END TIME 0:21

Digital Signature #
0D96E904E4704001
