

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2414

Name: RITTER 1  
Sterilizer ID: V2178198  
10/11/2022 09:58 AM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.5 F 0.0  
2:00 199.8 F 1.9  
4:00 199.3 F 7.1  
6:00 220.8 F 13.5  
8:00 232.2 F 9.1  
10:00 235.6 F 8.3  
12:00 243.2 F 11.2  
14:00 247.7 F 13.6  
16:00 261.5 F 21.4  
18:00 271.9 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.8  
1:00 271.8 F 27.8  
2:00 272.4 F 28.3  
3:00 272.8 F 28.7  
  
Min 271.4 F 27.7  
Max 272.8 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 229.1 F 0.0  
5:00 217.7 F 0.0  
10:00 212.4 F 0.0  
Dry Cycle Aborted

FILLING: 2:00  
HEATING: 18:07  
STERILIZING: 3:00  
VENTING: 2:45  
DRYING: 10:01  
TOTAL CYCLE: 00:35:53

10/11/2022 10:33 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_