
Midmark M11 - v1.0.5
Total Cycles: 3777

Name: RITTER2
Sterilizer ID: V2313228
07/12/2024 01:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.7 F 0.0
2:00 132.2 F 1.7
4:00 146.0 F 4.2
6:00 172.7 F 7.5
8:00 190.3 F 11.3
10:00 211.7 F 4.2
12:00 221.1 F 8.3
14:00 233.2 F 12.6
16:00 232.7 F 9.1
18:00 239.4 F 11.1
20:00 243.9 F 12.9
22:00 255.0 F 18.7
24:00 265.1 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 273.0 F 29.9
2:00 271.5 F 29.2
3:00 271.7 F 29.7
4:00 273.0 F 30.4

Min 271.0 F 29.1
Max 273.6 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 206.3 F 0.0
10:00 194.4 F 0.0
15:00 191.8 F 0.0
20:00 191.2 F 0.0
25:00 191.4 F 0.0
30:00 190.1 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 25:60
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:04:52
CYCLE COMPLETE

07/12/2024 02:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____