
Midmark M11 - v1.0.5
Total Cycles: 3826

Name: RITTER2
Sterilizer ID: V2313228
08/01/2024 08:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.2 F 0.0
2:00 70.3 F 0.0
4:00 75.6 F 0.0
6:00 99.8 F 0.1
8:00 152.3 F 0.2
10:00 177.3 F 1.7
12:00 192.8 F 4.9
14:00 206.3 F 8.1
16:00 220.4 F 11.7
18:00 224.6 F 5.9
20:00 233.4 F 10.1
22:00 238.1 F 9.4
24:00 241.2 F 12.0
26:00 239.3 F 10.4
28:00 251.6 F 16.4
30:00 262.0 F 22.5
32:00 271.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.5 F 29.2
2:00 271.9 F 29.4
3:00 272.0 F 28.9
4:00 272.6 F 29.6

Min 271.4 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 198.4 F 0.0
10:00 194.8 F 0.0
15:00 198.5 F 0.0
20:00 204.5 F 0.0
25:00 207.5 F 0.0
30:00 209.1 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:33
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:10:50
CYCLE COMPLETE

08/01/2024 09:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____