
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007646
9:18 10/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:25
STERILIZING 3:32
136.5 C 222kPa 3:32
136.2 C 224kPa 4:02
136.3 C 224kPa 4:32
135.8 C 219kPa 5:02
136.2 C 222kPa 5:32
136.4 C 223kPa 6:02
136.2 C 224kPa 6:32
136.4 C 222kPa 7:02
136.4 C 222kPa 7:02
VENTING 7:02
AIR DRYING 7:49
CYCLE COMPLETE 19:10

Digital Signature #
7165118AD75876D0
