

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3864

Name: RITTER2  
Sterilizer ID: V2313228  
08/15/2024 10:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 91.2 F 0.0  
2:00 91.3 F 0.0  
4:00 104.6 F 0.1  
6:00 139.1 F 1.9  
8:00 163.7 F 4.4  
10:00 172.4 F 7.4  
12:00 189.3 F 10.9  
14:00 214.6 F 6.1  
16:00 223.7 F 8.6  
18:00 233.8 F 12.9  
20:00 235.4 F 9.7  
22:00 238.5 F 9.9  
24:00 244.9 F 13.2  
26:00 255.0 F 18.9  
28:00 265.6 F 25.1  
30:00 270.6 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 30.1  
1:00 272.5 F 30.2  
2:00 272.5 F 30.5  
3:00 272.1 F 30.3  
4:00 272.6 F 30.2  
  
Min 271.9 F 30.0  
Max 273.0 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.3 F 0.0  
5:00 221.4 F 0.0  
10:00 222.4 F 0.0  
15:00 218.9 F 0.0  
20:00 224.0 F 0.0  
25:00 236.3 F 0.0  
30:00 231.5 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 30:55  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:09:31  
CYCLE COMPLETE

08/15/2024 11:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_