
Midmark M11 - v1.0.5
Total Cycles: 3122

Name: RITTER2
Sterilizer ID: V2313228
11/14/2023 11:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.6 F 0.0
2:00 125.2 F 0.3
4:00 127.7 F 2.1
6:00 156.8 F 5.2
8:00 175.1 F 9.2
10:00 206.1 F 13.5
12:00 219.3 F 7.4
14:00 232.7 F 12.2
16:00 232.3 F 9.1
18:00 239.1 F 10.5
20:00 244.8 F 13.4
22:00 256.6 F 19.6
24:00 267.0 F 26.2
26:00 273.2 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.9
1:00 272.8 F 30.2
2:00 272.0 F 30.3
3:00 272.4 F 30.0
4:00 272.5 F 30.5

Min 271.7 F 29.7
Max 273.1 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 216.5 F 0.0
10:00 229.2 F 0.0
15:00 230.3 F 0.0
20:00 230.5 F 0.0
25:00 228.0 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:18
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:04:43
CYCLE COMPLETE

11/14/2023 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____