
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002526
14:51 20/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:43
STERILIZING 3:46
136.3 C 220kPa 3:46
136.4 C 226kPa 4:17
136.4 C 226kPa 4:47
136.3 C 221kPa 5:17
136.1 C 223kPa 5:47
136.3 C 224kPa 6:17
136.2 C 224kPa 6:47
136.2 C 223kPa 7:17
136.2 C 223kPa 7:17
VENTING 7:17
AIR DRYING 8:08
CYCLE COMPLETE 15:25

Digital Signature #
63E4671C718F285C
