
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000839
14:39 05/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:10
STERILIZING 4:30
136.4 C 233kPa 4:30
136.4 C 233kPa 5:00
136.3 C 232kPa 5:30
136.4 C 233kPa 6:00
136.5 C 233kPa 6:30
136.3 C 232kPa 7:00
136.4 C 233kPa 7:30
136.4 C 233kPa 8:00
136.4 C 233kPa 8:00
VENTING 8:00
AIR DRYING 8:55
CYCLE COMPLETE 23:55

Digital Signature #
D02143B74F13CB6F
