
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000700
11:48 16/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:15
STERILIZING 3:17
136.5 C 228kPa 3:17
136.2 C 231kPa 3:47
136.4 C 233kPa 4:17
136.4 C 231kPa 4:47
136.4 C 232kPa 5:17
136.3 C 232kPa 5:47
135.9 C 230kPa 6:17
CYCLE ABORTED 6:44
STOP BUTTON PRESSED
NOT STERILE
END TIME 6:44

Digital Signature #
8E795F57E767D49D
