

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3316

Name: RITTER2  
Sterilizer ID: V2313228  
01/24/2024 01:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.1 F 0.0  
2:00 167.0 F 3.2  
4:00 191.7 F 7.9  
6:00 210.9 F 13.4  
8:00 225.2 F 8.8  
10:00 236.4 F 9.8  
12:00 243.6 F 13.2  
14:00 244.9 F 12.9  
16:00 258.5 F 20.4  
18:00 270.8 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 273.0 F 29.6  
2:00 272.5 F 29.6  
3:00 272.8 F 29.7  
4:00 272.6 F 29.7  
  
Min 270.8 F 28.5  
Max 273.2 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 172.5 F 0.0  
5:00 231.6 F 0.0  
10:00 198.9 F 0.0  
15:00 196.5 F 0.0  
20:00 193.7 F 0.0  
25:00 191.4 F 0.0  
30:00 189.9 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 18:38  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 00:57:41  
CYCLE COMPLETE

01/24/2024 02:08 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_