

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5357

Name: RITTER 1  
Sterilizer ID: V2178198  
02/24/2025 09:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.0 F 0.0  
2:00 75.2 F 0.2  
4:00 79.9 F 1.3  
6:00 115.4 F 3.6  
8:00 145.8 F 6.2  
10:00 166.4 F 9.0  
12:00 185.7 F 12.2  
14:00 207.7 F 2.6  
16:00 211.8 F 5.7  
18:00 222.8 F 9.0  
20:00 232.3 F 12.7  
22:00 232.2 F 7.9  
24:00 243.2 F 12.6  
26:00 239.8 F 10.3  
28:00 251.6 F 16.0  
30:00 261.8 F 22.0  
32:00 271.5 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.4 F 28.7  
2:00 271.8 F 28.7  
3:00 272.0 F 28.5  
4:00 272.3 F 29.0  
  
Min 271.4 F 28.2  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.8 F 0.0  
5:00 166.2 F 0.0  
10:00 179.5 F 0.0  
15:00 181.0 F 0.0  
20:00 182.1 F 0.0  
25:00 185.3 F 0.0  
30:00 231.6 F 0.0  
  
Drying Complete

FILLING: 0:49  
HEATING: 32:37  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:11:09  
CYCLE COMPLETE

02/24/2025 10:58 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_