

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001050  
11:37 29/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:10  
PRESSURIZING 2:17  
STERILIZING 3:16  
136.0 C 221kPa 3:16  
136.4 C 222kPa 3:46  
136.4 C 223kPa 4:16  
136.1 C 223kPa 4:46  
136.2 C 224kPa 5:16  
136.3 C 225kPa 5:46  
136.1 C 223kPa 6:16  
136.3 C 224kPa 6:46  
136.3 C 224kPa 6:46  
VENTING 6:46  
AIR DRYING 7:38  
CYCLE COMPLETE 22:39  
-----

Digital Signature #  
367F65503414475F  
-----