
Midmark M11 - v1.0.0
Total Cycles: 1818

Name: RITTER 1
Sterilizer ID: V2178198
04/13/2022 09:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.8 F 0.0
2:00 161.8 F 2.6
4:00 174.4 F 6.6
6:00 196.0 F 11.5
8:00 215.5 F 5.3
10:00 228.0 F 10.4
12:00 231.4 F 7.1
14:00 242.2 F 12.7
16:00 241.2 F 11.0
18:00 254.2 F 17.7
20:00 265.7 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.7 F 29.3
2:00 272.0 F 28.7
3:00 271.6 F 28.8
4:00 272.6 F 29.3

Min 271.3 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 220.6 F 0.0
10:00 231.9 F 0.0
15:00 231.2 F 0.0
20:00 227.6 F 0.0
25:00 230.4 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 21:44
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:51
CYCLE COMPLETE

04/13/2022 10:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____