
Midmark M11 - v1.0.5
Total Cycles: 1706

Name: RITTER2
Sterilizer ID: V2313228
11/10/2022 10:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.6 F 0.0
2:00 130.4 F 0.0
4:00 136.3 F 0.4
6:00 176.8 F 2.4
8:00 199.4 F 6.7
10:00 218.2 F 11.5
12:00 227.9 F 7.6
14:00 239.3 F 13.0
16:00 239.4 F 11.2
18:00 238.7 F 10.2
20:00 252.8 F 17.2
22:00 264.8 F 24.4
24:00 271.5 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.9 F 29.9
2:00 271.6 F 29.5
3:00 272.8 F 30.0
4:00 271.7 F 29.3

Min 271.5 F 29.3
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 207.0 F 0.0
10:00 210.5 F 0.0
15:00 221.9 F 0.0
20:00 221.0 F 0.0
25:00 229.8 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:38
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:02:42
CYCLE COMPLETE

11/10/2022 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____