
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007352
7:55 10/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 2:00
STERILIZING 3:10
136.4 C 220kPa 3:10
136.3 C 223kPa 3:40
136.4 C 222kPa 4:10
136.4 C 224kPa 4:40
136.4 C 222kPa 5:10
136.1 C 222kPa 5:40
136.4 C 223kPa 6:10
136.4 C 224kPa 6:40
136.4 C 224kPa 6:40
VENTING 6:40
AIR DRYING 7:28
CYCLE COMPLETE 37:28

Digital Signature #
4064AB7A40AD0D0A
