

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 003495  
11:58 20/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:02  
PRESSURIZING 2:06  
STERILIZING 3:02  
136.0 C 220kPa 3:02  
136.2 C 223kPa 3:32  
136.2 C 223kPa 4:02  
136.4 C 225kPa 4:32  
136.3 C 222kPa 5:02  
136.4 C 223kPa 5:32  
136.4 C 222kPa 6:02  
136.4 C 225kPa 6:32  
136.4 C 225kPa 6:32  
VENTING 6:32  
AIR DRYING 7:22  
CYCLE COMPLETE 7:27

-----  
Digital Signature #  
7CFA69B6D46027C5  
-----