
Midmark M11 - v1.0.5
Total Cycles: 2268

Name: RITTER2
Sterilizer ID: V2313228
04/19/2023 12:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.3 F 0.0
2:00 147.6 F 2.4
4:00 161.6 F 5.8
6:00 188.9 F 10.4
8:00 213.0 F 4.9
10:00 227.1 F 10.0
12:00 231.0 F 8.2
14:00 241.9 F 12.9
16:00 241.8 F 11.8
18:00 254.9 F 18.6
20:00 266.3 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.4 F 29.7
2:00 272.0 F 29.8
3:00 272.2 F 29.5
4:00 271.9 F 29.8

Min 271.5 F 29.2
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 213.5 F 0.0
10:00 229.8 F 0.0
15:00 233.3 F 0.0
20:00 227.4 F 0.0
25:00 223.4 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:34
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:60
CYCLE COMPLETE

04/19/2023 01:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____