
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.6uS / 1.6ppm
CYCLE NUMBER 000012
12:56 13/01/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:10
PRESSURIZING 6:18
STERILIZING 8:23
136.2 C 225kPa 8:23
136.3 C 226kPa 8:53
136.2 C 225kPa 9:23
136.4 C 228kPa 9:54
136.4 C 227kPa 10:24
136.3 C 226kPa 10:54
136.4 C 228kPa 11:24
136.4 C 227kPa 11:53
VENTING 11:54
AIR DRYING 13:10
CYCLE COMPLETE 43:10

Digital Signature #
9B092DDEAC95D76F
