

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4295

Name: RITTER2  
Sterilizer ID: V2313228  
01/24/2025 04:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 114.0 F 0.0  
2:00 115.4 F 1.1  
4:00 129.7 F 3.5  
6:00 154.0 F 6.8  
8:00 179.3 F 10.9  
10:00 211.7 F 4.2  
12:00 226.6 F 8.0  
14:00 238.2 F 12.4  
16:00 233.1 F 8.7  
18:00 241.4 F 11.8  
20:00 243.5 F 12.3  
22:00 255.1 F 18.4  
24:00 265.5 F 24.8  
26:00 271.1 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.5  
1:00 272.3 F 29.9  
2:00 272.9 F 29.8  
3:00 273.0 F 29.5  
4:00 272.4 F 29.4

Min 271.8 F 29.4  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
5:00 223.5 F 0.0  
10:00 187.0 F 0.0  
15:00 173.1 F 0.0  
20:00 171.0 F 0.0  
25:00 171.6 F 0.0  
30:00 170.6 F 0.0

Drying Complete

FILLING: 1:03  
HEATING: 26:31  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:05:10  
CYCLE COMPLETE

01/24/2025 05:54 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_