
Midmark M11 - v1.0.0
Total Cycles: 3775

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2023 02:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.7 F 0.0
2:00 179.8 F 3.0
4:00 201.0 F 7.6
6:00 218.0 F 12.4
8:00 227.6 F 7.8
10:00 239.6 F 13.0
12:00 239.2 F 10.6
14:00 236.3 F 8.8
16:00 250.6 F 15.3
18:00 261.7 F 21.6
20:00 272.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.4 F 28.7
2:00 272.4 F 29.1
3:00 271.3 F 28.5
4:00 272.3 F 29.0

Min 271.2 F 28.4
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.5 F 0.0
5:00 232.7 F 0.0
10:00 228.9 F 0.0
15:00 228.0 F 0.0
20:00 223.1 F 0.0
25:00 228.0 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:00:42
CYCLE COMPLETE

09/14/2023 03:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____