
Midmark M11 - v1.0.0
Total Cycles: 2927

Name: RITTER 1
Sterilizer ID: V2178198
02/22/2023 09:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.2 F 0.0
2:00 154.6 F 2.9
4:00 171.6 F 6.3
6:00 190.5 F 10.6
8:00 211.7 F 3.9
10:00 223.2 F 9.1
12:00 234.1 F 10.1
14:00 238.6 F 10.8
16:00 235.5 F 8.6
18:00 250.2 F 15.3
20:00 262.0 F 22.1
22:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 29.0
2:00 271.4 F 28.3
3:00 272.4 F 29.2
4:00 271.9 F 28.6

Min 271.3 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 204.5 F 0.0
10:00 224.5 F 0.0
15:00 221.3 F 0.0
20:00 223.5 F 0.0
25:00 229.3 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:38
CYCLE COMPLETE

02/22/2023 10:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____