
Midmark M11 - v1.0.0
Total Cycles: 3587

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2023 01:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.0 F 0.0
2:00 125.5 F 0.0
4:00 153.3 F 1.1
6:00 174.2 F 3.6
8:00 195.8 F 6.6
10:00 208.8 F 10.6
12:00 215.6 F 5.6
14:00 227.4 F 8.3
16:00 237.9 F 12.7
18:00 236.4 F 9.5
20:00 239.3 F 9.9
22:00 246.7 F 13.6
24:00 258.0 F 19.7
26:00 268.5 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.3 F 28.7
2:00 272.2 F 28.9
3:00 271.9 F 28.4
4:00 272.5 F 29.1

Min 271.5 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 196.5 F 0.0
10:00 206.0 F 0.0
15:00 205.7 F 0.0
20:00 205.3 F 0.0
25:00 205.2 F 0.0
30:00 203.2 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 27:12
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:05:33
CYCLE COMPLETE

07/26/2023 02:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____