
Midmark M11 - v1.0.5
Total Cycles: 3760

Name: RITTER2
Sterilizer ID: V2313228
07/08/2024 01:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.0 F 0.0
2:00 141.5 F 0.8
4:00 164.8 F 3.7
6:00 190.9 F 7.8
8:00 205.4 F 12.5
10:00 220.0 F 6.2
12:00 230.6 F 10.7
14:00 231.5 F 7.6
16:00 243.4 F 13.3
18:00 242.7 F 12.1
20:00 255.1 F 18.5
22:00 266.2 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.5 F 29.6
2:00 271.8 F 28.9
3:00 271.8 F 29.4
4:00 272.3 F 29.5

Min 271.1 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 210.9 F 0.0
10:00 196.6 F 0.0
15:00 192.2 F 0.0
20:00 189.1 F 0.0
25:00 186.9 F 0.0
30:00 187.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:02:01
CYCLE COMPLETE

07/08/2024 02:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____