
Midmark M11 - v1.0.5
Total Cycles: 4297

Name: RITTER2
Sterilizer ID: V2313228
01/27/2025 09:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 83.4 F 0.0
4:00 115.0 F 0.2
6:00 111.4 F 2.0
8:00 143.5 F 4.8
10:00 174.4 F 8.1
12:00 196.1 F 12.1
14:00 211.7 F 4.1
16:00 222.4 F 7.5
18:00 233.6 F 11.5
20:00 229.9 F 7.4
22:00 242.0 F 12.8
24:00 239.7 F 10.9
26:00 251.9 F 16.9
28:00 262.6 F 23.2
30:00 272.7 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.2 F 29.4
2:00 272.7 F 29.9
3:00 271.8 F 29.4
4:00 272.3 F 29.4

Min 271.7 F 29.1
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.2 F 0.0
5:00 205.6 F 0.0
10:00 199.7 F 0.0
15:00 205.5 F 0.0
20:00 208.7 F 0.0
25:00 209.9 F 0.0
30:00 211.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:20
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:08:36
CYCLE COMPLETE

01/27/2025 11:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____