
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
4.0uS / 2.5ppm
CYCLE NUMBER 000008
9:20 13/01/2020
WRAPPED
134 C for 6 min.
CYCLE START 0:00
CONDITIONING 4:10
PRESSURIZING 16:03
STERILIZING 17:44
136.4 C 227kPa 17:44
136.4 C 227kPa 18:15
136.4 C 227kPa 18:45
136.4 C 227kPa 19:15
136.4 C 227kPa 19:45
136.2 C 225kPa 20:15
136.4 C 227kPa 20:45
136.5 C 227kPa 21:15
136.4 C 227kPa 21:45
136.3 C 226kPa 22:15
136.4 C 228kPa 22:45
136.4 C 227kPa 23:15
136.4 C 227kPa 23:44
VENTING 23:44
AIR DRYING 24:57
CYCLE COMPLETE 54:57

Digital Signature #
3C8EECF9CAF940DD
